


Your Hospital Name  
Address  
City, State Zip  
(000) 000-0000

## Calibration Report

Acquired on February 05,2004 09:38 AM



### Counting Parameters

Isotope: Cs-137  
Gain: 2

Detector: Well  
Count Time (sec): 46

### Spectrum Statistics

Range (keV): 0 - 1520  
Peak Energy (keV): 661  
Peak Counts: 564  
Total Counts: 325868

Counts Outside Range: 6  
Peak Count Rate (cpm): 735  
Total Count Rate (cpm): 425045

### ROI Statistics

Energy Window(keV): 561-761  
Peak (keV): 661  
Peak Counts: 564

ROI Counts: 113665  
Peak cpm: 735  
ROI Rate (cpm): 148258

Technologist: John Smithson

Comments:

Signature: \_\_\_\_\_

Date : \_\_\_\_\_