


Background Results

Biodex Model 086-332 Serial # 08070015 Detector Serial # 8070003

03/03/2009 09:18
Measure Date



Full Spectrum

Counts: 447
Total CPM: 447.0
Test Time (secs): 60
Cnts >800 keV: 142

Status: **Ok** Technologist: **AMA**

	Isotope	Background Activity	
	Wide Window	303.0 CPM	879 DPM
1	Ba-133	56.0 CPM	455 DPM
2	Co-57	55.0 CPM	63 DPM
3	Co-58	239.0 CPM	1,972 DPM
4	Co-60	142.0 CPM	279 DPM
5	Cr-51	35.0 CPM	821 DPM
6	Cs-137	37.0 CPM	302 DPM
7	Ga-67	204.0 CPM	400 DPM
8	I-123	36.0 CPM	28 DPM
9	I-125	43.0 CPM	151 DPM
10	I-131	38.0 CPM	106 DPM
11	In-111	148.0 CPM	235 DPM
12	Sr-85	56.0 CPM	325 DPM
13	Tc-99m	35.0 CPM	45 DPM
14	Tl-201	65.0 CPM	131 DPM

Approved By: _____

Date: _____