


**Biodex Medical Systems, Inc**  
**20 Ramsay Road**  
**Shirley, NY 11967**

**Calibration - Wipe Counter**



Full Spectrum

Counts: 213074

Total CPM: 639,222.0

Test Time (secs): 20

Cnts >800 keV: 246

**Wipe Calibration  
Completed - Ok**

**FWHM = 5.9%**  
**(should be less than 10%)**

**Technologist**

**AMA**

Approved By: \_\_\_\_\_

Date: \_\_\_\_\_