

Stored Chi-Square Report

Biodex Model 086-332 Serial # 08070015 Detector Serial # 8070003

11/12/2008 11:56
Measure Date

Chi-Square Test Results

Value: 6.40 Status: Passed

Sample #	Counts
1	11722
2	11409
3	11530
4	11487
5	11552
6	11598
7	11539
8	11486
9	11628
10	11459

Chi-Square Table

Probability	Value
0.99	2.09
0.95	3.33
0.90	4.17 (Lower Limit)
0.50	8.34
0.10	14.68 (Upper Limit)
0.05	16.92
0.01	21.67

Technologist: AMA

Approved By: _____

Date: _____

Chi-Square Data

Biodex Model 086-332 Serial # 08070015 Detector Serial # 9020011
Printed: 04/27/2009 15:54

Print Range: All

Date	Detector SN	Status	Value	Technologist
03/27/09 13:04	0	PASS	7.32%	AMA
03/26/09 09:54	0	PASS	6.98%	AMA
03/13/09 13:13	0	PASS	9.00%	AMA
02/20/09 09:13	8070003	PASS	9.13%	AMA
02/20/09 09:09	8070003	PASS	9.27%	AMA
01/28/09 14:42	8070003	PASS	4.57%	AMA
01/28/09 14:39	8070003	PASS	9.50%	AMA

Approved By: _____

Date: _____